

## HVL evaluation of orthovoltage X-ray machine using EGSnrc code of simulation

A. Shabestani Monfared<sup>1</sup>, T. Allahverdi Pourfallah<sup>2\*</sup>, H. Babapour<sup>3</sup>,  
A.R. Shirazi<sup>4</sup>

<sup>1</sup>Biochemistry and Biophysics Department, Babol University of Medical Sciences, Shahid Rajaee Oncology Hospital, Babolsar, Iran

<sup>2</sup>Biochemistry and Biophysics Department, Faculty of Medicine, Mazandaran University of Medical Sciences, Sari, Iran

<sup>3</sup>Nuclear Engineering Department, Azad Islamic University- Sciences and Researches Branch, Tehran, Iran

<sup>4</sup>Medical Physics Department, Faculty of Medicine, Tehran University of Medical Sciences, Tehran, Iran

### ABSTRACT

#### ► Original article

**\* Corresponding author:**

Dr. Tayyeb Allahverdi Pourfallah,

Fax: +98 11 33543248

E-mail: [tpourfallah@mazums.ac.ir](mailto:tpourfallah@mazums.ac.ir)

Revised: June 2013

Accepted: Feb. 2014

*Int. J. Radiat. Res., October 2014;*  
12(4): 325-330

**Background:** Making use of the orthovoltage machines in Radiotherapy, is one of the routine methods for the treatment of the superficial lesions. In this study, an important determinant of X-ray quality, the HVL (Half Value Layer), has been evaluated. **Materials and Methods:** The HVLs of a orthovoltage X-ray machine in 120 and 180 kVp are measured, using an empirical method, in which the HVLs are derived from the absorption curves. The measured HVLs are compared with calculated (Monte Carlo simulation) HVLs. Using the BEAMDP code of simulation, the output spectra are obtained and employed for the measurement of the HVLs. **Results:** Comparing the calculated and measured HVL values, the results show that the highest and lowest differences between the two are 4.96% and 2.27%, respectively, which are, in fairly good agreement with those obtained in the former studies. **Conclusion:** This study shows that the EGSnrc simulation code is capable of being used for the extraction of the quality indices for the superficial X-ray radiotherapy machines. It seems that, the mentioned code, with the mentioned experimental method, can be employed as a routine clinical test tool for every superficial radiotherapy department.

**Keywords:** HVL, EGSnrc, MC simulation, orthovoltage machine.

### INTRODUCTION

Up to about 1950, most of the external radiotherapy was carried out with x-rays generated at voltages up to 300 kVp. Subsequent development of higher energy machines and the increasing popularity of the cobalt-60 units in the 1950s and the 1960s resulted in a gradual demise of the conventional kilovoltage machines. However, these machines have not been completely disappeared. Even in the present era of the megavoltage beams, there is some use of the lower energy beams, especially in the treatment of superficial lesions <sup>(1)</sup>.

Because X-ray beams used in superficial radiotherapy are always heterogeneous in energy, it is conventional sometimes to express the quality of an X-ray beam in terms of the effective energy. The effective energy of an X-ray beam is the energy of photons in a monoenergetic beam which is attenuated at the same rate as the radiation in question. In the case of low-energy X-ray beams (below megavoltage ranges), it is customary to describe quality in terms of HVL together with effective energy, although HVL (half-value layer) alone is adequate for most clinical application. HVL is the thickness of an absorber of specified composition required to

attenuate the intensity of the beam to half of its original value (1,2).

The availability of advanced Monte Carlo (MC) simulation systems and the ever increasing computing power has made Monte Carlo simulations of X-rays spectra an attractive addition and alternative to experimental measurements. As a consequence, the study and characterization of X-ray tubes and radiotherapy machines using Monte Carlo simulations has become common practice (3-7).

The applied MC codes in this study were the EGSnrc based BEAMnrc and BEAMDP from the NRCC group (8-10). The EGSnrc was developed from the EGS4 code by Nelson *et al.* (11). For this purpose the latest version of the EGSnrc code is used which includes directional bremsstrahlung splitting (DBS) to increase the efficiency of energy transition from the electron current to X-ray photons. The electron impact ionization model is also implemented which significantly improves the shape of the X-ray spectra (4, 12). The treatment head was simulated in the BEAMnrc code (9).

Determining and deriving the HVL and the output spectrum, in an orthovoltage machine at different kVps, using an empirical method and simulation (i.e., EGSnrc codes), is the purpose of this study. As the quality of orthovoltage beam is very important index in treating the superficial lesions and the quality is determined by HVL in these machines, deriving of this parameter could be very useful with practical aspects.

## MATERIALS AND METHODS

### Measurement method

For deriving the HVL of X-ray beam in the orthovoltage machine (Stubilipan, Siemens, Germany), exposure rates (air KERMA) were measured for 120 and 180kVp voltages at SSD=100. The experimental setup is schematically depicted in the figure 1. Machine specs were as follows: Energy range 120-300 kVp, tube current 12-20 mA, anode angle 30°, focal spot 8×8mm<sup>2</sup>, cone size 6×8cm<sup>2</sup>, and inherent filtration 2.4mmAl. Employed added filters for experiment were 0.2mmCu, 0.5mmCu and 1mm-

Cu. Tube current in these experiments was 18mA.

Farmer dosimeter (Nuclear Enterprise, US) is used for measuring the exposure rate, which is comprised of a thimble chamber and an electrometer. The dosimeter has been calibrated at SSDL (Secondary Standards Dosimetry Laboratories) of the IAEA (Iranian Atomic Energy Organization) to give air KERMA by considering correct quality factor.

The attenuation curves were drawn and the fitted equations were solved for deriving the HVLs (i.e., the point at which the exposure reaches to half of its original value) for 120 and 180kVp.

### Monte Carlo simulation

To model the orthovoltage machine, EGSnrcMP simulation code was used (13). The EGSnrc based MC user code BEAMnrc (9) was used to simulate the geometry of the head of the machine (i.e., including, source of the X-ray (X-ray tube), output window, inherent filter (2.4mm Al), collimator, added filter and applicator and outputs phase-space data (phase space files), which include all the particle information (i.e., the charge, position, direction, energy, and




Figure 1. Schematic geometry used for experiment (1).

history tag for each particle).

For simulating the source of the machine, the “parallel rectangular beam incident from side” with the following specs was selected (figure 2): target material Tungsten, anode angle 30°, focal spot size of 8×8mm<sup>2</sup>, cone size 6×8cm<sup>2</sup>, ECUT (electron energy cut off) 521 keV, PCUT (photons energy cut off) 1 keV.

The settings in EGSnrc were: Boundary crossing algorithm – EXACT, with skin depth 3 mean free path, electron steeping using PRESTA II, spin effects ON, simple bremsstrahlung angular sampling, and cross section for bremsstrahlung production according to NIST (National Institute of Standards and Technology). Bound Compton scattering, electron impact ionization and atomic relaxations were set to ON while photoelectron angular sampling was OFF. Geometry of simulation is depicted in figure 3.




Figure 2. X-ray source used in simulation (Rectangular type source with focal spot size of 8×8mm<sup>2</sup>).

**Determination of HVL using BEAMDP code**

For the determination of the HVL, phase space data at third scoring plane (i.e., output files of BEAMnrc code) were employed as input in the BEAMDP code. Then, the HVL was calculated, using the equation 1 that output file of the BEAMDP code had the required data for deriving the HVL (14).

The summation was performed over N energy bins with width ΔE<sub>i</sub> and midpoint energy

$$\frac{K(t)}{K(0)} = \left( \sum_{i=1}^N \varphi_i E_i \left\{ \frac{\mu_{en}}{\rho} (E_i) \right\}_{Air} \right) \cdot \exp[-\mu(E_i)_{Absorber} t] \Delta E_i \cdot \left( \sum_{i=1}^N \varphi_i E_i \left\{ \frac{\mu_{en}}{\rho} (E_i) \right\}_{Air} \Delta E_i \right)^{-1} = 0.5 \quad (1)$$




Figure 3. Geometry used for simulation. The scoring planes (1, 2, and 3) are also depicted.

E<sub>i</sub>. The photon fluence per energy bin φ<sub>i</sub> was extracted from the phase spaces with the utility program BEAMDP. The energy-dependent attenuation μ and mass absorption coefficients μ<sub>en</sub>/ρ are from Hubbell and Seltzer (15). This summation was performed iterative until a solution was found giving the thickness of the absorber to half the K in air.

In the simulation process, in each run, the variance reduction technique, DBS (directional bremsstrahlung splitting) (4), was also applied. This increases the efficiency of energy transition from the electron current to X-ray photons. The electron impact ionization model is also implemented, which significantly improves the shape of the X-ray spectra as described by Kawrakow (16).

The BEAMnrc code ran under the Fedora 7 Linux OS with a Pentium® 4 computer with 2×3 GHz CPU and 1Gbyte RAM. The numbers of histories in each run was 1.5×10<sup>8</sup> with the total CPU time 12.5h.

The EGSnrc based MC user code BEAMDP was used to derive the X-ray spectra at different energies (15, 17). For calculating the mean energy of the spectra the following equation was used (18);

$$E_{ave} = \frac{\sum_{i=1}^n E_i \Phi_i \Delta E_i}{\sum_{i=1}^n \Phi_i \Delta E_i} \quad (2)$$

Where,  $E_i$  is the bin energy,  $\Phi_i$  bin fluence,  $\Delta E_i$  bin difference between two consecutive energy bins.

### RESULTS

Figure 4 (a and b) shows the absorption curves for 120 and 180 kVp beam energies. In




Figure 4. Semi-log attenuation curves obtained using empirical method for 120 (a) and 180kVp (b) energies.

Table 1. HVL of X-ray beam in 120 and 180kVp obtained using experiment and simulation. Percentage difference between two methods,  $E_{eff}$  and  $E_{ave}$  are also shown.

Energy (kVp)	HVL(mmCu) (Experiment)	HVL(mmCu) (MC)	Difference (%)	$E_{eff}$ (keV)	$E_{ave}$ (keV)
120	0.264±0.046	0.272±6.86%	2.94%	49.9	60.9
180	0.479±0.017	0.504±5.45%	4.96%	57.2	79.4

these figures the mean exposure rates (nC/min) were drawn against the thickness of attenuator layers. The average absolute error of measurement for 120 and 180 kVp beam energies were ±0.032 and ±0.017, respectively (n=3).

Table 1 shows the results of measurement and simulation in deriving the HVL for different beam energies. The effective energy ( $E_{eff}$ ) and average energy ( $E_{ave}$ ) of the beam are also shown in the table. Percentage difference between simulation and measurement were at most 4.96%.

The figure 5 (a and b) show the output spectra for 180 and 120 kVp beam energies. These beam spectra were drawn at specified scoring planes using the BEAMDP code. The phase space data at these scoring planes were obtained using the BEAMnrc code.




Figure 5. Output spectra for orthovoltage X-ray in 180 (a) and 120kVp (b) energies. Average percentage relative errors of simulations were 4.91 and 5.45% respectively.

## DISCUSSION

In this study, the thicknesses of HVL layers, are calculated, using EGSnrc simulation codes, and measured, using an empirical method at different energies (120 and 180 kVp), then, are compared with each other. In the empirical method, the absorption curves were employed for measuring the HVL. At first, these curves were drawn, by fitting a third grade equation over the points at which the transmitted exposures through the attenuation layers have been measured.

By interpolation over the drawn curves, the HVLs (i.e., points at which the intensity of the exposure reached to half of its original value) are derived at different energies.

The EGSnrc based BEAMnrc and DOSXYZnrc from the National Research Council of Canada (NRCC) group were used in this study<sup>(19)</sup>. The EGSnrc was developed from the EGS4 code<sup>(11)</sup>. The EGS (Electron-Gamma-Shower) system of computer codes is a general purpose package for the Monte Carlo simulation of the coupled transport of electrons and photons in an arbitrary geometry for particles with energies above a few keV up to several hundreds of GeV<sup>(13)</sup>. The EGSnrc-based BEAMnrc code<sup>(20, 21)</sup> was used to simulate the geometry of the head of the machine and the BEAMDP code was used to derive the X-ray spectra at different energies and determining of the HVL<sup>(4, 12)</sup>.

In the previous studies the HVL index was evaluated using different experimental methods and also simulation codes (e.g., PENELOPE). Mainegra *et al.*<sup>(4)</sup> proposed two methods for calculating the HVL, that were in good agreement with experimental results (i.e., 2.3% differences at most). Chica *et al.*<sup>(3)</sup> in their study calculated the HVL by the PENELOPE code, which showed about 2% difference with the experiment. Another method for the evaluation of the X-ray beam quality was proposed by Chica *et al.*<sup>(3)</sup> in which the ratio of absorbed dose in  $Z_{1/2}$  and  $Z_{1/4}$  depths (which are the depths in water at which the dose is 50 and 25% of the dose at  $Z_0$ ) were measured and compared with those obtained by simulation (PENELOPE). From the simulated and measured depth dose

curves they have evaluated the corresponding  $Z_{1/2}$  and  $Z_{1/4}$ . As for the quality indices, no statistical differences between experimental and simulated results are observed, and both the  $Z_{1/2}$  and the  $Z_{1/4}$  obtained from the experimental curves are in reasonable good agreement with the simulated ones. The comparison between simulated and experimental results showed a very good agreement (at most 3%).

The practical worth of this study is applicability of this method in radiotherapy departments with limited lab instruments. In this study, comparison between the calculated and measured HVL values (table 1) showed that the highest and lowest differences between the two are 4.96% and 2.27%, respectively, which are in fairly good agreement with those obtained in the former studies. These differences were lower than those obtained in other studies<sup>(22, 23)</sup>.

## CONCLUSION

The results of this study showed the capability of EGSnrc simulation code in extracting the quality indices of superficial X-ray radiotherapy machines. This study showed that this code and the mentioned empirical method can be employed as a routine clinical test tool for every radiotherapy department, especially in those with limited lab instruments.

**Conflict of interest:** Declared none.

## REFERENCES

1. Johns HE, Cunningham JR (1983) The Physics of Radiology. 4th edn. Springfield & Illinois: Charles C. Thomas Publisher.
2. Khan FM (2003) Physics of Radiation Therapy. 3rd edn. Philadelphia: Lippincott Williams & Wilkins.
3. Chica U, Anguiano M, Lallena AM (2009) Benchmark of penelope for low and medium energy X-rays. *Phys Med*, **25**: 51-57.
4. Mainegra-Hing E, Kawrakow I (2006) Efficient X-ray tube simulations. *Med Phys*, **33**: 2683-2690.
5. Allahverdi Pourfallah T, Allahverdi M, Riahi Alam N, Ay M, Zahmatkesh M (2009) Verifying the accuracy of dose distribution in gamma knife unit in presence of inhomogeneities using PAGAT polymer gel dosimeter and MC simulation. *Iran J Radiat Res*, **7**: 49-56.

6. Pourfallah T, Allahverdi M, Alam N, Ay M, Zahmatkesh M, Ibbott G (2009) Performance evaluation of MRI-based PAGAT polymer gel dosimeter in an inhomogeneous phantom using EGSnrc code on a Co-60 machine. *Appl Radiat Isot*, **67**: 186-191.
7. Pourfallah T, Allahverdi M, Riahi Alam N, Ay M, Zahmatkesh M (2009) Differential dose volume histograms of Gamma Knife in the presence of inhomogeneities using MRI-polymer gel dosimetry and MC simulation. *Med Phys*, **36**: 3002-3012.
8. Rogers DWO, Faddegon BA, Ding GX, Ma CM, Wei J, Mackie TR (1995) BEAM: A Monte Carlo code to simulate radiotherapy treatment units. *Med Phys*, **22**: 503 – 524.
9. Rogers D, Walters B, Kawrakow I (2007) BEAMnrc Users Manual. Ottawa : National Research Council of Canada; 2007 May. 268 p. Report PIRS-0509(A) revL.
10. Ma CM and Rogers DWO (2009) BEAMDP Users Manual : National Research Council of Canada; 2009 July. 35 p. Report PIRS-0509(C) revA.
11. Nelson W, Hirayama H, Rogers D (1985) The EGS4 Code System. Stanford Linear Accelerator Center, Stanford, California; 1985. Report SLAC-265.
12. Kawrakow I and Rogers DWO (2003) The EGSnrc code system: Monte Carlo simulation of electron and photon transport. Technical Report PIRS-701, 4th printing: Ottawa, Canada: National Research Council of Canada.
13. Kawrakow I, Mainegra-Hing E, Rogers D, Tessier F, Walters B (2011) The EGSnrc Code System: Monte Carlo Simulation of Electron and Photon Transport. Ottawa: National Research Council of Canada; 2011. Report PIRS-701.
14. Knöös T, Rosenschöld PM, Wieslander E (2007) Modelling of an Orthovoltage X-ray Therapy Unit with the EGSnrc Monte Carlo Package Radiation Physics. *J Phys, Conf. Ser.:* 021009.
15. Hubbell JH and Seltzer SM (2004) Tables of X-ray mass attenuation coefficients and mass energy-absorption coefficients from 1 keV to 20 MeV for elements Z=1 to 92 and 48 additional substances of dosimetric interest. National Institute of Standards and Technology, Gaithersburg: NISTIR 5632. Version 5631.5634 [Online].
16. Kawrakow I (2002) Electron impact ionization cross sections for EGSnrc. *Med Phys*, **29**: 1230.
17. Berger MJ and Hubbell JH (1987) XCOM: Photon Cross Sections on a Personal Computer. NBSIR 87-3597, National Bureau of Standards (former name of NIST), Gaithersburg, MD.
18. Mainegra-Hing E and Kawrakow I (2006) Efficient X-ray tube simulations. *Med Phys*, **33**: 2683-2690.
19. Rogers DWO, Faddegon BA, Ding GX, Ma CM, Wei J, Mackie TR (1995) BEAM: A Monte Carlo code to simulate radiotherapy treatment units. *Med Phys*, **22**: 503 – 524.
20. Nelson WR, Hirayama H, Rogers DWO (1985) The EGS4 Code System. Stanford Linear Accelerator Center, Stanford, California; 1985. Report SLAC-265.
21. Treurniet J, Walters B, Kawrakow I, Rogers D (2005) BEAMnrc, DOSXYZnrc and BEAMDP GUI User's Manual. Ottawa: National Research Council of Canada; 2005 Jul. 17 p. Report PIRS-0623(rev C).
22. Verhaegen F, NA E, Van de Putte S, Namito Y (1999) Monte Carlo modelling of radiotherapy kV X-ray units. *Phys Med Biol*, **44**: 1767-1789.
23. Omrane LB, Verhaegen F, Chahed N, Mtimet S (2003) An investigation of entrance surface dose calculations for diagnostic radiology using Monte Carlo simulations and radiotherapy dosimetry formalisms. *Phys Med Biol*, **48**: 1809-1824.