

# Dosimetric comparison of the heart substructures with IMRT and VMAT techniques in left breast radiotherapy: The effect of deep inspiratory breath-hold

A. Arslan<sup>1\*</sup>, E. Aktaş<sup>2</sup>, S.K. Eren<sup>3</sup>, I. Dengiz<sup>4</sup>, S.A. Arslan<sup>5</sup>, Y. Güneş<sup>5</sup>

<sup>1</sup>Kayseri City Hospital, Clinic of Radiation Oncology, Doctor, Kayseri City Hospital, Kayseri, Turkey

<sup>2</sup>Kayseri City Hospital, Clinic of Radiology, Doctor, Kayseri City Hospital, Kayseri, Turkey

<sup>3</sup>Kayseri City Hospital, Clinic of General Surgery, Doctor, Kayseri City Hospital, Kayseri, Turkey

<sup>4</sup>Kayseri City Hospital, Clinic of Radiation Oncology, Health Physicist, Kayseri City Hospital, Kayseri, Turkey

<sup>5</sup>Ankara Memorial Hospital, Clinic of Radiation Oncology, Doctor, Ankara Memorial Hospital, Ankara, Turkey

## ABSTRACT

### ► Original article

#### \*Corresponding author:

Alaettin Arslan, Ph.D.,

#### E-mail:

alaettin.arslan@gmail.com

Received: February 2022

Final revised: August 2022

Accepted: August 2022

Int. J. Radiat. Res., January 2023;  
21(1): 23-30

DOI: 10.52547/ijrr.21.1.3

**Background:** This study aimed to compare the doses received by the four chambers and vascular structures of the heart during adjuvant radiotherapy (RT) after left breast-conserving surgery (BCS) using intensity-modulated RT (IMRT) and volumetric-modulated arc therapy (VMAT) techniques. **Material and Methods:** Simulation images were taken of 14 patients who underwent left BCS with both free-breathing (FB) and deep inspiration breath-hold (DIBH) techniques. Left breast RT was planned with both IMRT and VMAT. Planned target volumes 50 and 60, homogeneity index, conformity index, and monitor unit values, as well as radiation doses received by organs at risk, were compared. **Results:** In IMRT compared to VMAT, in the heart ( $D_{mean}$ ,  $V_{10}$ ,  $V_2$ ) and heart substructures (left ventricle [ $V_5$ ], right ventricle [ $D_{mean}$ ], right atrium [ $D_{mean}$ ,  $D_{max}$ ], left atrium [ $D_{mean}$ ,  $V_5$ ], right coronary artery [RCA;  $D_{mean}$ ,  $D_{max}$ ], left artery coronary main [LACM;  $D_{mean}$ ], and left circumflex artery [LCXA;  $D_{mean}$ ,  $D_{max}$ ]), significant dose reductions were observed. When FB and DIBH results were compared, in the DIBH technique, the heart ( $D_{mean}$ ,  $V_{25}$ ,  $V_{10}$ ,  $V_2$ ) and heart substructures (left ventricle [ $D_{mean}$ ,  $D_{max}$ ,  $V_{23}$ ,  $V_5$ ], right ventricle [ $D_{mean}$ ,  $D_{max}$ ], right atrium [ $D_{mean}$ ,  $D_{max}$ ], left atrium [ $D_{mean}$ ,  $D_{max}$ ], left anterior descending artery [ $D_{mean}$ ,  $D_{max}$ ,  $V_{20}$ ], RCA [ $D_{max}$ ], LACM [ $D_{mean}$ ], and LCXA [ $D_{mean}$ ,  $D_{max}$ ]), doses were significantly decreased. **Conclusion:** In RT of patients with left BCS, significant dose reductions occurred in the lung, heart, and almost all substructures of the heart using DIBH compared to FB.

**Keywords:** Left breast cancer, breast-conserving surgery, IMRT, VMAT, DIBH, heart substructure radiation doses.

## INTRODUCTION

Adjuvant radiotherapy (RT) in breast cancer carries long-term cardiac risks. In particular, conventional planning techniques and large irradiation areas increase this risk <sup>(1)</sup>. The relative risk of cardiac mortality is estimated to increase by 0.04 per Gy of radiation to the heart <sup>(2)</sup>. Two hypotheses regarding RT-induced cardiovascular damage have been proposed. The first hypothesis stated that radiation increases the frequency of myocardial infarction by interacting with one or more steps of the pathogenic pathway of age-related coronary artery atherosclerosis. The second hypothesis proposed that the lethality of myocardial infarction increases due to pathologies unrelated to radiation <sup>(3)</sup>. It has been stated that cardiac risk increases, particularly in the first 5 years after treatment, and continues for 20 years <sup>(4)</sup>.

Cardiac risk can be minimized by reducing the doses taken by the heart and the heart's substructures during breast RT. It has been shown that contouring the heart's substructures as organs at

risk (OARs) can reduce the dose to these structures without compromising the dose distribution, even in conventional and free-breathing (FB) techniques <sup>(5)</sup>. In addition, dose reduction in OARs has become more pronounced with the introduction of advanced technologies such as intensity-modulated RT (IMRT), volumetric-modulated arc therapy (VMAT), deep inspiration breath-hold (DIBH), and proton therapy <sup>(6-8)</sup>. Breast RT with the DIBH technique was first described in a study published in 2001, and significant reductions in the radiation doses received by the heart were shown with this technique <sup>(9)</sup>. In the following years, breast RT with DIBH has attracted significant interest, and many studies have been published on its efficacy and dose-reduction capacity <sup>(10-14)</sup>. In ongoing studies, the effects of the DIBH technique, together with different planning modalities, on the ipsilateral lung, total lung, heart, left ventricle and left anterior descending artery (LAD) are being investigated.

In our study, we examined the effects of RT plans on the other 3 chambers of the heart, right coronary artery (RCA), left artery coronary main (LACM), and

left circumflex artery (LCxA), in addition to other known OARs. We aimed to compare the doses received by OARs during FB and DIBH techniques employed during IMRT and VMAT plans.

## MATERIALS AND METHODS

### Patient selection and planning

Between January 2020 and June 2021, 14 patients who underwent left breast-conserving surgery (BCS) and were referred for RT in our clinic were included in the study. Two simulation images with 3 mm sections were taken of the patients, using both FB and DIBH techniques. The Philips Brilliance Big Bore Computed Tomography Simulator was used to take the images and the RPM Respiratory Gating System (version 1.7.5; Varian Medical Systems) was used for respiratory monitoring. OARs were contoured on the images under the guidance of a radiologist with reference to the Radiation Therapy Oncology Group guidelines and heart atlas study by Feng *et al.* (15). While the left breast was contoured and a dose of 50 Gy was defined, a dose of 60 Gy was defined for the tumor bed by contouring, with the support of a general surgeon, the lumpectomy tumor site, clips, and seroma; 95% of the target volumes (D95%) were intended to receive treatment doses of 50 and 60 Gy with a 95%–107% dose homogeneity. The simultaneous integrated boost technique and 6 MV X-rays in 28 fractions (178.5 cGy/day for the left breast, 214 cGy/day for the tumor bed) were preferred for dose administration. Plans with 7–9 fields in FB and 5–7 fields in DIBH were made for IMRT, while plans with 4 half arcs were made in FB and DIBH for VMAT. All plans were made using the Eclipse Treatment Planning System (version 15.1; Varian Medical Systems). The Varian Clinac IX was used as the RT treatment device. The radiation dose received by the planned target volume (PTV) 50 and 60, homogeneity index (HI), conformity index (CI), monitor unit (MU), and OARs (both lungs, right breast, esophagus, spinal cord, whole heart, and substructures of the heart) were compared for the 4 plans made for each patient. Dose distributions and dose-volume histograms of the same patient in 4 different techniques are shown in figures 1 and 2, respectively.


### Statistical analysis

The conformity of the data to the normal distribution was evaluated by histogram, Q-Q plots, and the Shapiro–Wilk test. Homogeneity of variance was tested with Levene's test. The paired *t*-test was applied for quantitative measurements with 2 replicates. Data were evaluated with R 4.0.0 software (www.r-project.org). The significance level was accepted as  $p < 0.05$ .


### Power analysis

A power analysis is performed to identify the

necessary sample size. For  $\alpha = 0.05$ , power = 0.85, and effect size = 0.788, the minimum sample size was 14. Power analyses were conducted using GPower 3.1.9.7 software.



**Figure 1.** Dose distribution image for patient X. 95% of the target volumes (D95%) were intended to receive a 50 and 60 Gy treatment dose, with a dose homogeneity of 95–107%. **a** IMRT-FB Intensity-modulated radiotherapy-Free breath, **b** IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, **c** VMAT-FB Volumetric-modulated arc therapy-Free breath, **d** VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.



**Figure 2.** Dose-volume histogram (DVH) image for patient X. The colored curves in DVH represent: Dark green PTV 50, Red PTV 60, Pink Heart, Brown Left ventricle, Orange Right ventricle, Blue LAD, Dark blue RCA, Cyan LACM, Magenta LCxA, and Yellow Left lung. PTV Planning target volume, LAD Left anterior descending artery, RCA Right coronary artery, LACM Left artery coronary main, LCxA Left circumflex artery. **a** IMRT-FB Intensity-modulated radiotherapy-Free breath, **b** IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, **c** VMAT-FB Volumetric-modulated arc therapy-Free breath, **d** VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.

## RESULTS

Table 1 compares PTV, HI, CI, and MU values for the 4 planning techniques. Table 2 shows the doses

received by the OARs during IMRT and VMAT. Table 3 presents the doses received by the OARs when the FB and DIBH techniques were employed.

**Table 1.** Comparison of PTV, HI, CI, and MU values.

Parameters		IMRT FB	IMRT DIBH	<i>p</i>	VMAT FB	VMAT DIBH	<i>p</i>
PTV 60	D%2(Gy)	63.68±0.66	63.66±0.64	0.902	63.21±0.60	63.20±0.45	0.945
	D%98(Gy)	59.79±0.44	59.62±0.54	0.335	59.73±0.44	59.84±0.31	0.126
	D%50(Gy)	61.93±0.40	62.03±0.32	0.315	61.96±0.36	61.92±0.28	0.679
PTV 50	D%98(Gy)	49.35±1.05	49.15±0.64	0.498	48.59±0.50	48.61±0.54	0.827
	D%50(Gy)	52.94±0.53	53.16±0.43	0.059	52.81±0.49	52.80±0.68	0.922
HI		0.06±0.01	0.06±0.02	0.876	0.05±0.01	0.05±0.01	0.435
CI		1.19±0.08	1.21±0.07	0.129	1.11±0.03	1.11±0.05	0.870
MU		1827.79±462.62	1487.64±425.59	<b>&lt;0.001</b>	532.36±42.47	521.93±41.62	0.279

Values are expressed as mean±SD. The bold value indicates statistical significance (*p* <0.05). IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric -modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold, PTV Planning target volume, HI Homogeneity index, (ICRU 83/HI of zero is ideal); (D2%-D98%) /D50%, CI Conformity index, (ICRU 62/CI of 1.0 is ideal); Volume of PTV covered by the 95% isodose curve/volume of PTV, MU monitor unit.

**Table 2.** Comparison of OAR doses at IMRT and VMAT.

Parameters		IMRT FB	VMAT FB	<i>p</i>	IMRT DIBH	VMAT DIBH	<i>p</i>
Heart	<b>D<sub>mean</sub> (Gy)</b>	8.59±1.78	9.33±1.59	<b>0.049</b>	5.74±1.58	7.27±1.51	0.362
	<b>V<sub>25</sub> (%)</b>	5.56±3.26	4.15±2.62	<b>0.006</b>	2.16±2.06	1.33±1.68	<b>0.017</b>
	<b>V<sub>10</sub> (%)</b>	22.00±10.16	28.80±9.81	<b>0.040</b>	11.53±8.24	17.29±8.56	<b>&lt;0.001</b>
	<b>V<sub>2</sub> (%)</b>	98.83±3.43	99.89±0.40	<b>&lt;0.001</b>	89.79±10.03	98.68±4.31	<b>0.002</b>
Ipsilateral Lung	<b>D<sub>mean</sub> (Gy)</b>	12.72±2.32	13.25±1.42	<b>0.018</b>	12.03±2.25	12.79±1.94	<b>0.035</b>
	<b>V<sub>20</sub> (%)</b>	18.25±4.70	18.99±2.87	0.414	17.57±4.60	18.81±4.17	0.085
	<b>V<sub>12</sub> (%)</b>	32.30±9.99	35.38±6.09	0.191	28.92±7.35	34.31±7.65	<b>0.001</b>
	<b>V<sub>5</sub> (%)</b>	69.87±17.50	85.41±10.40	<b>&lt;0.001</b>	61.19±14.53	79.15±13.07	<b>&lt;0.001</b>
Esophagus	<b>D<sub>mean</sub> (Gy)</b>	5.22±4.37	6.82±3.55	<b>0.002</b>	4.59±3.80	6.36±3.28	<b>0.001</b>
Spinal Cord	<b>D<sub>max</sub> (Gy)</b>	11.09±10.54	9.92±7.30	0.325	7.87±8.62	9.16±7.92	0.079
LAD	<b>D<sub>mean</sub> (Gy)</b>	17.90±8.47	15.27±7.85	<b>0.004</b>	11.40±5.51	10.98±3.80	0.603
	<b>D<sub>max</sub> (Gy)</b>	38.50±13.76	32.65±11.72	<b>0.004</b>	31.67±16.68	25.17±11.66	<b>0.011</b>
	<b>V<sub>40</sub> (%)</b>	14.51±18.92	4.21±9.99	<b>0.047</b>	4.67±8.73	0.56±2.11	0.053
	<b>V<sub>20</sub> (%)</b>	24.22±22.92	28.96±27.18	0.370	14.14±17.39	11.46±14.22	0.317
Left Ventricle	<b>D<sub>mean</sub> (Gy)</b>	10.90±3.74	9.31±2.73	<b>0.001</b>	6.72±2.56	7.10±2.74	0.394
	<b>D<sub>max</sub> (Gy)</b>	45.47±12.42	40.08±11.14	<b>0.018</b>	34.67±14.93	29.65±13.00	0.063
	<b>V<sub>23</sub> (%)</b>	10.33±8.73	5.94±6.57	<b>0.005</b>	3.48±3.93	1.42±2.09	<b>0.025</b>
	<b>V<sub>5</sub> (%)</b>	84.06±19.60	84.26±16.97	0.951	49.68±30.80	62.31±25.42	<b>0.023</b>
Left Atrium	<b>D<sub>mean</sub> (Gy)</b>	3.75±0.91	4.56±0.95	<b>0.021</b>	3.04±1.28	4.59±1.06	<b>0.001</b>
	<b>D<sub>max</sub> (Gy)</b>	7.37±2.18	7.00±1.78	0.601	5.83±2.23	7.59±2.95	0.079
	<b>V<sub>5</sub> (%)</b>	16.00±21.86	32.23±34.72	0.154	9.02±17.80	27.16±30.32	<b>0.049</b>
Right Ventricle	<b>D<sub>mean</sub> (Gy)</b>	9.95±2.92	11.15±2.45	0.091	6.74±3.03	8.20±3.19	<b>0.001</b>
	<b>D<sub>max</sub> (Gy)</b>	43.34±10.16	35.70±8.56	<b>&lt;0.001</b>	30.33±14.74	25.40±9.26	<b>&lt;0.001</b>
Right Atrium	<b>D<sub>mean</sub> (Gy)</b>	3.77±1.03	6.67±2.25	<b>&lt;0.001</b>	2.83±0.98	5.10±1.88	<b>0.001</b>
	<b>D<sub>max</sub> (Gy)</b>	8.93±4.73	15.41±6.53	<b>0.001</b>	6.41±3.57	11.50±4.34	<b>&lt;0.001</b>
Right Breast	<b>D<sub>mean</sub> (Gy)</b>	3.29±0.82	5.85±0.95	<b>&lt;0.001</b>	2.58±0.87	5.92±1.04	<b>&lt;0.001</b>
Right Lung	<b>D<sub>mean</sub> (Gy)</b>	4.60±2.43	7.07±1.58	<b>&lt;0.001</b>	3.56±1.52	6.54±0.79	<b>&lt;0.001</b>
	<b>V<sub>5</sub> (%)</b>	28.04±19.01	60.62±15.48	<b>&lt;0.001</b>	19.65±14.83	53.84±10.27	<b>&lt;0.001</b>
Total Lung	<b>D<sub>mean</sub> (Gy)</b>	8.28±2.14	9.84±1.34	<b>0.001</b>	7.46±1.60	9.37±1.14	<b>&lt;0.001</b>
	<b>V<sub>20</sub> (%)</b>	9.31±3.85	9.78±2.14	0.421	8.63±2.65	9.65±2.07	<b>0.018</b>
RCA	<b>D<sub>mean</sub> (Gy)</b>	6.67±2.63	11.30±3.85	<b>&lt;0.001</b>	4.93±1.87	9.28±3.16	<b>0.002</b>
	<b>D<sub>max</sub> (Gy)</b>	8.51±3.45	15.20±4.97	<b>&lt;0.001</b>	6.59±2.56	12.00±3.82	<b>0.002</b>
LACM	<b>D<sub>mean</sub> (Gy)</b>	5.27±1.51	5.34±1.02	<b>0.005</b>	4.32±2.18	5.73±1.24	<b>0.026</b>
	<b>D<sub>max</sub> (Gy)</b>	6.02±1.93	5.87±1.36	0.793	4.84±2.45	6.12±1.29	0.057
LCxA	<b>D<sub>mean</sub> (Gy)</b>	5.97±1.69	5.51±0.92	0.383	4.48±1.48	5.68±1.10	<b>0.004</b>
	<b>D<sub>max</sub> (Gy)</b>	7.28±2.33	6.56±1.27	0.322	5.63±2.15	6.75±1.34	<b>0.036</b>

Values are expressed as mean±SD. The bold value indicates statistical significance (*p* <0.05). IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric -modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold, D<sub>mean</sub> mean dose, D<sub>max</sub> maximum dose, V<sub>X</sub> (%) The percent volume of organ receiving X Gy dose, LAD Left anterior descending artery, RCA Right coronary artery, LACM Left artery coronary main, LCxA Left circumflex artery.

Table 3. Comparison of OAR doses at FB and DIBH.

Parameters		IMRT FB	IMRT DIBH	<i>p</i>	VMAT FB	VMAT DIBH	<i>p</i>
Heart	<i>D<sub>mean</sub></i> (Gy)	8.59±1.78	5.74±1.58	0.390	9.33±1.59	7.27±1.51	<0.001
	<i>V<sub>25</sub></i> (%)	5.56±3.26	2.16±2.06	<0.001	4.15±2.62	1.33±1.68	<0.001
	<i>V<sub>10</sub></i> (%)	22.00±10.16	11.53±8.24	<0.001	28.80±9.81	17.29±8.56	<0.001
	<i>V<sub>2</sub></i> (%)	98.83±3.43	89.79±10.03	<0.001	99.89±0.40	98.68±4.31	0.266
Ipsilateral Lung	<i>D<sub>mean</sub></i> (Gy)	12.72±2.32	12.03±2.25	0.018	13.25±1.42	12.79±1.94	0.192
	<i>V<sub>20</sub></i> (%)	18.25±4.70	17.57±4.60	0.221	18.99±2.87	18.81±4.17	0.817
	<i>V<sub>12</sub></i> (%)	32.30±9.99	28.92±7.35	0.038	35.38±6.09	34.31±7.65	0.527
	<i>V<sub>5</sub></i> (%)	69.87±17.50	61.19±14.53	0.023	85.41±10.40	79.15±13.07	0.014
Esophagus	<i>D<sub>mean</sub></i> (Gy)	5.22±4.37	4.59±3.80	0.262	6.82±3.55	6.36±3.28	0.356
Spinal Cord	<i>D<sub>max</sub></i> (Gy)	11.09±10.54	7.87±8.62	0.034	9.92±7.30	9.16±7.92	0.193
LAD	<i>D<sub>mean</sub></i> (Gy)	17.90±8.47	11.40±5.51	0.001	15.27±7.85	10.98±3.80	0.021
	<i>D<sub>max</sub></i> (Gy)	38.50±13.76	31.67±16.68	0.081	32.65±11.72	25.17±11.66	0.011
	<i>V<sub>40</sub></i> (%)	14.51±18.91	4.67±8.73	0.058	4.21±9.99	0.56±2.11	0.211
	<i>V<sub>20</sub></i> (%)	24.22±22.92	14.14±17.39	0.076	28.96±27.18	11.46±14.22	0.006
Left Ventricle	<i>D<sub>mean</sub></i> (Gy)	10.90±3.74	6.72±2.56	<0.001	9.31±2.72	7.10±2.74	0.001
	<i>D<sub>max</sub></i> (Gy)	45.47±12.42	34.67±14.93	0.002	40.08±11.14	29.65±13.00	<0.001
	<i>V<sub>23</sub></i> (%)	10.33±8.73	3.48±3.93	0.006	5.94±6.57	1.42±2.09	0.009
	<i>V<sub>5</sub></i> (%)	84.06±19.60	49.68±30.80	<0.001	84.26±16.97	62.31±25.42	<0.001
Left Atrium	<i>D<sub>mean</sub></i> (Gy)	3.75±0.91	3.04±1.28	0.037	4.56±0.95	4.59±1.06	0.912
	<i>D<sub>max</sub></i> (Gy)	7.37±2.18	5.83±2.23	0.002	7.00±1.78	7.59±2.95	0.471
	<i>V<sub>5</sub></i> (%)	16.00±21.86	9.02±17.80	0.229	32.23±34.72	27.16±30.32	0.548
Right Ventricle	<i>D<sub>mean</sub></i> (Gy)	9.95±2.92	6.74±3.03	0.001	11.15±2.45	8.20±3.19	0.001
	<i>D<sub>max</sub></i> (Gy)	43.34±10.16	30.33±14.74	0.001	35.70±8.56	25.40±9.26	<0.001
Right Atrium	<i>D<sub>mean</sub></i> (Gy)	3.77±1.03	2.83±0.98	0.004	6.67±2.25	5.10±1.88	0.014
	<i>D<sub>max</sub></i> (Gy)	8.93±4.73	6.41±3.57	0.013	15.41±6.53	11.50±4.34	0.032
Right Breast	<i>D<sub>mean</sub></i> (Gy)	3.29±0.82	2.58±0.87	0.007	5.85±0.95	5.92±1.04	0.653
Right Lung	<i>D<sub>mean</sub></i> (Gy)	4.60±2.43	3.56±1.52	0.181	7.07±1.58	6.54±0.79	0.176
	<i>V<sub>5</sub></i> (%)	28.04±19.01	19.65±14.83	0.190	60.62±15.48	53.84±10.27	0.027
Total Lung	<i>D<sub>mean</sub></i> (Gy)	8.28±2.14	7.46±1.60	0.067	9.84±1.34	9.37±1.14	0.117
	<i>V<sub>20</sub></i> (%)	9.31±3.85	8.63±2.65	0.420	9.78±2.14	9.65±2.07	0.771
RCA	<i>D<sub>mean</sub></i> (Gy)	6.67±2.63	4.93±1.87	0.063	11.30±3.85	9.28±3.16	0.062
LACM	<i>D<sub>max</sub></i> (Gy)	8.51±3.45	6.59±2.56	0.134	15.20±4.97	12.00±3.82	0.031
	<i>D<sub>mean</sub></i> (Gy)	5.27±1.51	4.32±2.18	0.002	5.34±1.02	5.73±1.24	0.298
	<i>D<sub>max</sub></i> (Gy)	6.02±1.93	4.84±2.45	0.109	5.87±1.36	6.12±1.29	0.578
LCxA	<i>D<sub>mean</sub></i> (Gy)	5.97±1.69	4.48±1.48	<0.001	5.51±0.92	5.68±1.10	0.673
	<i>D<sub>max</sub></i> (Gy)	7.28±2.33	5.63±2.15	0.001	6.56±1.27	6.75±1.34	0.700

Values are expressed as mean±SD. The bold value indicates statistical significance (*p* <0.05). IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric-modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold, *D<sub>mean</sub>* mean dose, *D<sub>max</sub>* maximum dose, *V<sub>X</sub>* (%) The percent volume of organ receiving X Gy dose, LAD Left anterior descending artery, RCA Right coronary artery, LACM Left artery coronary main, LCxA Left circumflex artery.

### PTV coverage, HI, CI, and MU

There was no difference between the VMAT and IMRT plans in PTV 60 D2% (near-maximum dose), D98% (near-minimum dose), and D50% (mean dose) values or PTV 50 D98% and D50% values. There was no significant difference between VMAT and IMRT in HI and CI values, though the MU value was significantly lower in IMRT-DIBH compared to IMRT-FB (*p* < 0.001; table 1).

### Heart dose

The lowest *D<sub>mean</sub>* dose to the heart, 5.74 Gy, occurred with the IMRT-DIBH plan, while the highest, 9.33 Gy, occurred with the VMAT-FB plan. The *V<sub>25</sub>* volume was lowest, at 1.33%, with the VMAT-DIBH plan and highest, at 5.56%, with the IMRT-FB plan. *V<sub>10</sub>* volume was lowest with IMRT-DIBH (11.53%) and highest with VMAT-FB (28.80%), and *V<sub>2</sub>* volume was lowest with IMRT-DIBH (89.79%) and highest with VMAT-FB (99.89%; figure 3, tables 2 and 3).

### Heart substructure dose

**Left ventricle:** The lowest *D<sub>mean</sub>* dose to the left

ventricle, 6.72 Gy, occurred with the IMRT-DIBH plan, and the highest, 10.90 Gy, occurred with the IMRT-FB plan. The lowest *D<sub>max</sub>* dose occurred with VMAT-DIBH (29.65 Gy), while the highest occurred with IMRT-FB (45.47 Gy). *V<sub>23</sub>* volume was lowest with VMAT-DIBH (1.42%) and highest with IMRT-FB (10.33%). The lowest *V<sub>5</sub>* volume was seen with IMRT-DIBH (49.68%) and the highest was seen with VMAT-FB (84.26%; figure 4, tables 2 and 3).

**Right ventricle:** The lowest *D<sub>mean</sub>* dose to the right ventricle, 6.74 Gy, occurred with the IMRT-DIBH plan, while the highest, 11.15 Gy, occurred with the VMAT-FB plan. The lowest *D<sub>max</sub>* dose occurred with VMAT-DIBH (25.40 Gy) and the highest occurred with IMRT-FB (43.34 Gy; tables 2 and 3).

**Left atrium:** The lowest *D<sub>mean</sub>* dose to the left atrium, 3.04 Gy, was provided by the IMRT-DIBH plan, and the highest, 4.59 Gy, was provided by the VMAT-DIBH plan. The lowest *D<sub>max</sub>* dose was 5.83 Gy and occurred with the IMRT-DIBH plan, while the highest was 7.59 Gy and occurred with the VMAT-DIBH plan. The lowest *V<sub>5</sub>* volume was seen with the IMRT-DIBH plan





(9.02%) and the highest was seen with the VMAT-FB plan (32.32%; tables 2 and 3).

**Right atrium:** The lowest  $D_{mean}$  dose to the right atrium was 2.83 Gy and occurred with the IMRT-DIBH plan, while the highest was 6.67 Gy and occurred with the VMAT-FB plan. The lowest  $D_{max}$  dose, 6.41 Gy, was provided by the IMRT-DIBH plan, while the highest, 15.41 Gy, was provided by the VMAT-FB plan (tables 2 and 3).

**LAD:** The lowest  $D_{mean}$  dose to the LAD, 10.98 Gy,


occurred with the VMAT-DIBH plan, and the highest, 17.90 Gy, occurred with the IMRT-FB plan. The lowest  $D_{max}$  dose of 25.17 Gy occurred with VMAT-DIBH, while the highest  $D_{max}$  dose of 38.50 Gy occurred with IMRT-FB.  $V_{40}$  volume was lowest with VMAT-DIBH (0.56%) and highest with IMRT-FB (14.51%), and  $V_{20}$  volume was lowest with VMAT-DIBH (11.46%) and highest with VMAT-FB (28.96%; figure 5, tables 2 and 3).

**Figure 3.** Mean $\pm$ 1.96\*Standard error. Graphical view of heart Dmean (a), V25 (b), V10 (c) and V2 (d) values in four different planning techniques. Dmean mean dose, VX (%) The percent volume of organ receiving X Gy dose IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric-modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.



**Figure 4.** Mean $\pm$ 1.96\*Standard error. Graphical view of left ventricle Dmean (a), Dmax (b), V23 (c) and V5 (d) values in four different planning techniques. Dmean mean dose, Dmax maximum dose, VX (%) The percent volume of organ receiving X Gy dose, IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric-modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.

**Figure 5.** Mean $\pm$ 1.96\*Standard error. Graphical view of LAD Dmean (a), Dmax (b), V40 (c) and V20 (d) values in four different planning techniques. Dmean mean dose, Dmax maximum dose, VX (%) The percent volume of organ receiving X Gy dose, LAD Left anterior descending artery, IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric-modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.



**RCA:** The lowest  $D_{mean}$  dose to the RCA was 4.93 Gy and occurred with the IMRT-DIBH plan, and the highest was 11.30 Gy and occurred with the VMAT-FB plan. The lowest  $D_{max}$  dose, 6.59 Gy, occurred with IMRT-DIBH, and the highest, 15.20 Gy, occurred with VMAT-FB (tables 2 and 3).


**LACM:** The lowest  $D_{mean}$  dose to the LACM, 4.32 Gy, was seen with the IMRT-DIBH plan, and the highest, 5.73 Gy, was seen with the VMAT-DIBH plan. The lowest  $D_{max}$  dose was 4.84 Gy and occurred with the IMRT-DIBH plan, while the highest was 6.12 Gy and occurred with the VMAT-DIBH plan (tables 2 and 3).

**LCxA:** The lowest  $D_{mean}$  dose to the LCxA, 4.48 Gy, occurred with the IMRT-DIBH plan, and the highest, 5.97 Gy, occurred with the IMRT-FB plan. The lowest

$D_{max}$  dose of 5.63 Gy occurred with IMRT-DIBH, and the highest  $D_{max}$  dose of 7.28 Gy occurred with IMRT-FB (tables 2 and 3).

#### Ipsilateral lung dose

The lowest  $D_{mean}$  dose to the ipsilateral lung was 12.03 Gy and provided by the IMRT-DIBH plan, while the highest was 13.25 Gy and provided by the VMAT-FB plan.  $V_{20}$  volume was lowest with the IMRT-DIBH plan (17.57%) and highest with the VMAT-FB plan (18.99%).  $V_{12}$  volume was lowest with IMRT-DIBH (28.92%) and highest with VMAT-FB (35.38%), and  $V_5$  volume was lowest with IMRT-DIBH (61.19%) and highest with VMAT-FB (85.41%; figure 6, tables 2 and 3).



**Figure 6.** Mean $\pm$ 1.96\*Standard error. Graphical view of ipsilateral (left) lung Dmean (a), V20 (b), V12 (c) and V5 (d) values in four different planning techniques. Dmean mean dose, VX (%) The percent volume of organ receiving X Gy dose, IMRT-FB Intensity-modulated radiotherapy-Free breath, IMRT-DIBH Intensity-modulated radiotherapy-Deep inspiration breath-hold, VMAT-FB Volumetric-modulated arc therapy-Free breath, VMAT-DIBH Volumetric-modulated arc therapy-Deep inspiration breath-hold.



### Other OARs

A comparison of the right lung, total lung, right breast, esophagus and spinal cord doses are given in tables 2 and 3 in detail.

## DISCUSSION

RT for breast cancer is an important component of disease management and can reduce the absolute risk of breast cancer mortality, though it can cause serious late side effects in the OARs (i.e., heart and lungs). The effect of DIBH on the heart, LAD, and ipsilateral lung has been demonstrated in previous studies reporting that this technique is superior to the FB technique<sup>(16-18)</sup>. Darby et al. found a linear relationship between the heart  $D_{mean}$  and the rate of major coronary events, which increased by 7.4% per Gy of heart  $D_{mean}$ <sup>(1)</sup>. In the study by Mathieu et al., a decrease of approximately 3 times in the heart  $D_{mean}$  and approximately 3.5 times in the LAD  $D_{mean}$  was obtained with the DIBH technique compared to FB<sup>(19)</sup>. In Gaál et al.'s study, while significant decreases were observed in the heart  $D_{mean}$  and  $V_{25}$  (%), LAD  $D_{mean}$  and  $D_{max}$ , and ipsilateral lung  $D_{mean}$  and  $V_{20}$  (%) values with the DIBH technique compared to the FB technique ( $p < 0.001$ ), a slight increase was observed in the right breast dose, and the HI value was similar for both techniques<sup>(20)</sup>. Ferdinand et al. prescribed a 40 Gy hypofractionated scheme in 15 fractions with preferred electron treatment to the tumor bed and found a reduction in the heart  $D_{mean}$  dose from 4 Gy to 2.4 Gy and the heart  $V_{10}$  from 8.9% to 3.4% with DIBH compared to FB. In the same study, LAD  $D_{mean}$  was reduced from 12.6 Gy to 8.7 Gy, LAD  $D_{max}$  was reduced from 31.9 Gy to 25.8 Gy, and LAD  $V_{40}$  was reduced from 0.6% to 0.4%<sup>(21)</sup>. In the study by Yu et al., FB and DIBH techniques were compared in IMRT and VMAT plans. There was no significant difference between heart and LAD results using DIBH. However, while VMAT was found to have significantly lower ipsilateral lung  $V_{30}$  (%), IMRT had significantly lower right lung  $D_2$  (Gy), right breast  $D_2$  (Gy), and right breast  $V_5$  (%). VMAT-DIBH provided much lower doses than VMAT-FB to almost all OARs, which is in line with the results of our study<sup>(22)</sup>. Zhang et al. compared FB and DIBH techniques in VMAT and found that heart, LAD, left and right lung, and right breast doses were significantly lower in the DIBH technique<sup>(23)</sup>. In our study, we observed the lowest heart  $D_{mean}$ ,  $V_{10}$ , and  $V_2$  values with the IMRT-DIBH technique but the lowest  $V_{25}$  value with the VMAT-DIBH technique. Although post-RT cardiac side effects have been primarily described in relation to heart  $D_{mean}$  and tangential fields, the anterior apical portion of the heart is most likely to receive the highest doses<sup>(24)</sup>. As dose distribution in the heart is not homogeneous, the radiation received by substructures of the heart could be altered with modern techniques. Jacob et al. demonstrated that heart  $D_{mean}$  was insufficient to predict left ventricle and LAD doses; hence, the doses received by these substructures are necessary when

evaluating cardiotoxicity<sup>(25)</sup>. We also observed inconsistency between heart  $D_{mean}$ , left ventricle  $D_{mean}$ , and LAD  $D_{mean}$ , as heart  $D_{mean}$  and left ventricle  $D_{mean}$  were lowest with the IMRT-DIBH technique but LAD  $D_{mean}$  was lowest with VMAT-DIBH. The IMRT-DIBH technique was also superior for the right atrium and right ventricle. In the current study, regardless of the plan employed, a significant reduction in the amount of radiation received by the heart was achieved when using DIBH compared to FB.

In a recent meta-analysis by Taylor et al., 10 years after RT, the risk of radiation-related lung cancer increased by approximately 11% (95% confidence interval 6–19) per Gy of mean lung dose<sup>(2)</sup>. Pneumonitis is another possible complication of breast cancer RT that can lead to lung fibrosis several months after treatment<sup>(26)</sup>. As with radiation-related lung cancer, the risk of pneumonitis increases with increasing lung radiation dose. In this context, while an intermediate dose such as  $V_{20}$  is a well-established risk factor for radiation pneumonitis,  $V_5$ , often caused by IMRT and VMAT, may also be associated with pneumonitis<sup>(27)</sup>.

In our study, we observed the lowest doses in all parameters for the ipsilateral lung with IMRT-DIBH. Moreover, DIBH significantly reduced ipsilateral and contralateral lung  $V_5$  in both IMRT and VMAT techniques. Radiation-induced coronary artery disease is characterized by ostial stenosis with a significantly higher incidence of severe LACM disease, followed by ostial RCA and LAD stenosis. The location and severity of stenosis directly correlate with the volume irradiated and the dose of the radiation beam<sup>(28)</sup>. A few studies in the literature have evaluated the RCA, LACM, and LCxA in breast cancer RT<sup>(29-31)</sup>. However, no comparisons have been made regarding how these vascular structures are affected by the use of FB and DIBH techniques in IMRT and VMAT plans. In the current study, lower doses were observed in all 3 of these vascular structures with IMRT-DIBH. In the DIBH technique, IMRT compared to VMAT, all results except for LACM  $D_{max}$  were significantly lower. In the IMRT plans, DIBH elicited lower values than FB in all outcomes, with significance present in LACM  $D_{mean}$  and LCxA  $D_{mean}$  and  $D_{max}$ . In VMAT plans, the RCA dose was lower in the DIBH technique than in the FB technique, while the LACM and LCxA doses were higher in the DIBH technique than the FB technique. Long-term follow-up is required to understand how these differences in coronary artery dose and volume affect the clinical outcome.

Of the limitations of this study, the most obvious is the challenge of contouring the vasculature of the heart. Although relevant atlases and the aid of a radiologist were employed to optimize the contouring, these vascular structures are occasionally seen only faintly on computed tomography sections. Using a 0.5 cm brush pen in the planning system, the arteries were contoured in the sections where they were visible, and automatic joining was used in the other sections.

## CONCLUSION

In RT of patients with left BCS, while IMRT and VMAT provided variable advantages and disadvantages in the doses received by OARs, the DIBH technique provided significant dose reductions in the lung, heart, and all substructures of the heart compared to FB. Based on these results, we recommend using the DIBH technique in addition to the patient-specific IMRT or VMAT plan in RT of patients with left BCS.

**Conflict Of Interests:** The authors have no conflict of interest to declare.

**Funding Sources:** No financial contribution of any kind has supported this publication.

**Ethical approval:** All procedures performed in studies involving human participants were in accordance with the ethical standards of the institutional and/or national research committee and with the 1964 Helsinki Declaration and its later amendments or comparable ethical standards.

**Informed consent:** Institutional review board approval was obtained for this study. The study was conducted with the approval of the Non-Interventional Clinical Research Ethics Committee of Kayseri City Hospital. Ethics Committee convened on 11.06.2020 and received the protocol number 2020/78.

**Authors' Contribution:** All authors had full access to the data in the study and took responsibility for the integrity of the data and the accuracy of the data analysis. *Conceptualization*, AA, SAA, EA, ID; *Methodology*, AA, SAA, EA, SKE, ID; *Investigation*, AA, EA, SKE, ID, YG; *Formal Analysis*, AA, ID, YG; *Writing-Original Draft, Review & Editing*, AA; *Visualization*, AA.

## REFERENCES

- Darby SC, Ewertz M, McGale P, Bennet AM, Blom-Goldman U, et al. (2013) Risk of ischemic heart disease in women after radiotherapy for breast cancer. *New England Journal of Medicine*, **368**(11): 987-998.
- Taylor C, Correa C, Duane FK, Aznar MC, Anderson SJ, et al. (2017) Early Breast Cancer Trialists' Collaborative Group. Estimating the risks of breast cancer radiotherapy: evidence from modern radiation doses to the lungs and heart and from previous randomized trials. *Journal of Clinical Oncology*, **35**(15): 1641.
- Andratschke N, Maurer J, Molls M, Trott KR (2011) Late radiation-induced heart disease after radiotherapy. Clinical importance, radiobiological mechanisms and strategies of prevention. *Radiotherapy and Oncology*, **100**(2): 160-166.
- Henson KE, McGale P, Taylor C, Darby SC (2013) Radiation-related mortality from heart disease and lung cancer more than 20 years after radiotherapy for breast cancer. *British Journal of Cancer*, **108**(1): 179-182.
- Arslan A, Aktas E, Sengul B, Tekin B (2021) Dosimetric evaluation of left ventricle and left anterior descending artery in left breast radiotherapy. *La Radiologia Medica*, **126**(1): 14-21.
- Loap P, Fourquet A, Kirova Y (2020) The limits of the linear quadratic (LQ) model for late cardiotoxicity prediction: example of hypofractionated rotational intensity modulated radiation therapy (IMRT) for breast cancer. *Int. J Radiat Oncol Biol Phys*, **106**(5): 1106-1108.
- Lauche O, Kirova YM, Fenoglietto P, Costa E, Lemanski C, et al. (2016) Helical tomotherapy and volumetric modulated arc therapy: new therapeutic arms in the breast cancer radiotherapy. *World Journal of Radiology*, **8**(8): 735.
- Fagundes M, Hug EB, Pankuch M, Fang C, McNeeley S, et al. (2015) Proton therapy for locally-regionally advanced breast cancer maximizes cardiac sparing. *Int. J Particle Therapy*, **1**(4): 827-844.
- Sixel KE, Aznar MC, Ung YC (2001) Deep inspiration breath hold to reduce irradiated heart volume in breast cancer patients. *Int. J Radiat Oncol Biol Phys*, **49**(1): 199-204.
- Bartlett FR, Colgan RM, Carr K, Donovan EM, McNair HA, et al. (2013). The UK HeartSpare Study: randomised evaluation of voluntary deep-inspiratory breath-hold in women undergoing breast radiotherapy. *Radiotherapy and Oncology*, **108**(2): 242-247.
- Latty D, Stuart KE, Wang W, Ahern V (2015) Review of deep inspiration breath-hold techniques for the treatment of breast cancer. *Journal of Medical Radiation Sciences*, **62**(1): 74-81.
- Rice L, Goldsmith C, Green MM, Cleator S, Price PM (2017) An effective deep-inspiration breath-hold radiotherapy technique for left-breast cancer: impact of post-mastectomy treatment, nodal coverage, and dose schedule on organs at risk. *Breast Cancer: Targets and Therapy*, **9**: 437.
- Dell'Oro M, Giles E, Sharkey A, Borg M, Connell C, et al. (2019) A retrospective dosimetric study of radiotherapy patients with left-sided breast cancer; patient selection criteria for deep inspiration breath hold technique. *Cancers*, **11**(2): 259.
- Lai J, Hu S, Luo Y, Zheng R, Zhu Q, et al. (2020) Meta-analysis of deep inspiration breath hold (DIBH) versus free breathing (FB) in postoperative radiotherapy for left-side breast cancer. *Breast Cancer*, **27**(2): 299-307.
- Feng M, Moran JM, Koelling T, Chughtai A, Chan JL, et al. (2011) Development and validation of a heart atlas to study cardiac exposure to radiation following treatment for breast cancer. *Int J Radiat Oncol Biol Phys*, **79**(1): 10-18.
- Park S, Rim CH, Yoon WS (2021) Variation of heart and lung radiation doses according to setup uncertainty in left breast cancer. *Radiation Oncology*, **16**(1): 1-9.
- Chang CS, Chen CH, Liu KC, Ho CS, Chen MF (2020) Selection of patients with left breast cancer for IMRT with deep inspiration breath-hold technique. *Journal of Radiation Research*, **61**(3): 431-439.
- Song J, Tang T, Caudrelier JM, Bélec J, Chan J, Lacasse P, ..., Nair V (2021). Dose-sparing effect of deep inspiration breath hold technique on coronary artery and left ventricle segments in treatment of breast cancer. *Radiotherapy and Oncology*, **154**: 101-109.
- Mathieu D, Bedwani S, Mascolo-Fortin J, Côté N, Bernard A A, et al. (2020) Cardiac sparing with personalized treatment planning for early-stage left breast cancer. *Cureus*, **12**(3).
- Gaál S, Káhn Z, Paczona V, Kószó R, Drencsényi R, Szabó J, et al. (2021) Deep-inspirational breath-hold (DIBH) technique in left-sided breast cancer: various aspects of clinical utility. *Radiation Oncology*, **16**(1): 1-11.
- Ferdinand S, Mondal M, Mallik S, Goswami J, Das S, Manir KS, et al. (2021) Dosimetric analysis of deep inspiratory breath-hold technique (DIBH) in left-sided breast cancer radiotherapy and evaluation of pre-treatment predictors of cardiac doses for guiding patient selection for DIBH. *Technical Innovations & Patient Support in Radiation Oncology*, **17**: 25-31.
- Yu PC, Wu CJ, Tsai YL, Shaw S, Sung SY, Lui LT, Nien HH (2018) Dosimetric analysis of tangent-based volumetric modulated arc therapy with deep inspiration breath-hold technique for left breast cancer patients. *Radiation Oncology*, **13**(1): 1-10.
- Zhang W, Li R, You D, Su Y, Dong W, Ma Z (2020) Dosimetry and feasibility studies of volumetric modulated Arc therapy with deep inspiration breath-hold using optical surface management system for left-sided breast cancer patients. *Frontiers in Oncology*, **10**: 1711.
- Moignier A, Broggio D, Derreumaux S, El Baf F, Mandin AM, Girinsky T, et al. (2014) Dependence of coronary 3-dimensional dose maps on coronary topologies and beam set in breast radiation therapy: a study based on CT angiographies. *Int J Radiat Oncol Biol Phys*, **89**(1): 182-190.
- Jacob S, Camilleri J, Derreumaux S, Walker V, Lairez O, Lapeyre M, et al. (2019) Is mean heart dose a relevant surrogate parameter of left ventricle and coronary arteries exposure during breast cancer radiotherapy: a dosimetric evaluation based on individually-determined radiation dose (BACCARAT study). *Radiation Oncology*, **14**(1): 1-10.
- Marks LB (2002) Dosimetric predictors of radiation-induced lung injury. *Int J Radiat Oncol Biol Phys*, **54**(2): 313-316.
- Tsujino K, Hashimoto T, Shimada T, Yoden E, Fujii O, Ota Y, et al. (2014) Combined analysis of V20, V55, pulmonary fibrosis score on baseline computed tomography, and patient age improves prediction of severe radiation pneumonitis after concurrent chemoradiotherapy for locally advanced non-small-cell lung cancer. *Journal of Thoracic Oncology*, **9**(7): 983-990.
- Yang EH, Marmagkiolis K, Balanescu DV, Hakeem A, Donisan T, Finch W, et al. (2021) Radiation-induced vascular disease—a state-of-the-art review. *Frontiers in Cardiovascular Medicine*, **8**: 223.
- Gocer GPS and Ozer EE (2020). Effect of radiotherapy on coronary arteries and heart in breast-conserving surgery: a dosimetric analysis. *Radiology and Oncology*, **54**(1): 128.
- Naimi Z, Moujahed R, Neji H, Yahyaoui J, Hamdoun A, Bohli M, Kochbati L (2021) Cardiac substructures exposure in left-sided breast cancer radiotherapy: Is the mean heart dose a reliable predictor of cardiac toxicity?. *Cancer/Radiothérapie*, **25**(3): 229-236.
- Kataria T, Bisht SS, Gupta D, Abhishek A, Basu T, Narang K, et al. (2016) Quantification of coronary artery motion and internal risk volume from ECG gated radiotherapy planning scans. *Radiotherapy and Oncology*, **121**(1): 59-63.